



# CERTIFICATE OF ACCREDITATION

**The ANSI National Accreditation Board**

Hereby attests that

**Covert Manufacturing, Inc.**  
**328 South East Street**  
**Galion, OH 44833**

Fulfills the requirements of

**ISO/IEC 17025:2017**

In the field of

**DIMENSIONAL MEASUREMENT**

This certificate is valid only when accompanied by a current scope of accreditation document.  
The current scope of accreditation can be verified at [www.anab.org](http://www.anab.org).

A handwritten signature in black ink, appearing to read 'R. Douglas Leonard Jr.', is positioned above a horizontal line.

R. Douglas Leonard Jr., VP, PILR SBU

Expiry Date: 02 April 2024  
Certificate Number: L2042-1



This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017.  
This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory  
quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).

**SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017**

**Covert Manufacturing, Inc.**

328 South East Street  
Galion, OH 44833  
Jared Baker  
419-468-1761

**DIMENSIONAL MEASUREMENT**

Valid to: **April 2, 2024**

Certificate Number: **L2042-1**

**3 Dimensional**

Parameter	Range	Expanded Uncertainty of Measurement (+/-) <sup>2</sup>	Reference Standard, Method, and/or Equipment
Dimensional Measurement 3D	X = (0 to 1 524) mm Y = (0 to 3 048) mm Z = (0 to 1 120) mm	(8.5 + 0.035L) μm	Zeiss CMM DB1200 Utilized as Reference Standard for Dimensional Inspection
	(0 to 2 600) mm	(98 + 0.016L) μm	Faro Arm – (Non-Contact Scan Arm)

Calibration and Measurement Capability (CMC) is expressed in terms of the measurement parameter, measurement range, expanded uncertainty of measurement and reference standard, method, and/or equipment. The expanded uncertainty of measurement is expressed as the standard uncertainty of the measurement multiplied by a coverage factor of 2 ( $k=2$ ), corresponding to a confidence level of approximately 95%.

Notes:

1. On-site service is available for this parameter, since on-site conditions are typically more variable than those in the laboratory, larger measurement uncertainties are expected on-site than what is reported on the accredited scope.
2.  $L$  = Length in millimeters.
3. This scope is formatted as part of a single document including Certificate of Accreditation No. L2042-1.



R. Douglas Leonard Jr., VP, PILR SBU